

Surat Municipal Institute of Medical Education and Research

Opp. Bombay Market, Umarwada, Surat – 395 010 (Guj.)-India.

Medical Superintendent-Extn.1600
SMIMER Hospital
Phone Nos. 2368040 to 43

FAX No.(0261) 2360306



SMIMER/Hosp/Out/ No.6838

Date: 27 / 03 /2025

To,
M/s _____

Sub: Quotation for repairing/replacement the Various parts of Various Instruments of Medicine,Surgery & Emergency department of SMIMER Hospital.

Dear Sir,

Quotations are invited for item as per attached list and as per following terms & conditions.

1. The rates should be F.O.R. destination inclusive of all taxes & charges i.e. GST etc.
 2. Rate should be quoted as per serial number of list.
 3. The rejected items shall have to be removed by the supplier at their own cost.
 4. Overwriting, erasures or correction in quotations shall not be considered.
 5. Quotations are to be sent in sealed cover as above.
 6. Quotations received after due date will not be considered.
 7. Quotations sent by post or courier services is preferable.
 8. Guarantee/warranty should be mentioned in quotation.
 9. Supplier has to submit quality certificate & samples if requires.
 10. Effective goods will not be accepted.
- GST Clause For Construction/ Erection/ Commissioning/Installation /Repairs/ Maintenance/Renovation/ Fabrication Of Structure Including Building(Means All Works Contract/Turn Key Projects/Supply Of Material/Goods)
 - GST (Goods & Service Tax) has come in existence from 1st July, 2017. Contractor/ Successful Bidder is bound to pay any amount of GST prescribed by the Govt. of India as per the Terms of contract agreed upon during the course of execution of this Contract.

- During the course of execution of contract, if there is any change in Rate of GST(Goods & Service Tax) by the Government, the same shall be reimbursed / recovered separately by SMC, subject to **the submission of original Receipt/Proof for the amounts actually remitted by** the successful Tenderer /Contractor to the Competent Authority along with a Certificate from Chartered Accountant of Contractor / Successful Bidder certifying that the amount of GST paid to the Government and the same shall be intimated / submitted/claimed within 15 (fifteen) Days from the date of payment. Remittance of GST within stipulated period shall be the sole responsibility of the Successful Bidder/Contractor, failing which, SMC may recover the amount due, from any other payable dues with SMC and decision of Municipal Commissioner shall be final and binding on the Contractor / Successful Bidder in this regard. Further, the non-payment of GST to the Government may lead to the termination of the contract and forfeiture of Security Deposit / Performance Guarantee Amount.
- If imposition of any other new Taxes /Duties /Levies / Cess or any other incidentals etc. or any increase in the existing Taxes / Duties / Levies /Cess or any other incidentals etc. (Excluding GST) are imposed during the course of the contract, the same shall be borne by the Contractor / Successful Bidder only, in no case SMC shall be liable for the same.
- **The contractor will submit the invoice to the SMC having GSTIN of SMC mentioned therein and the taxes shall be shown separately on the face of the invoice so as to claim as ITC by SMC.**

Envelope should bear on top: " Quotation for repairing/replacement the Various parts of Various Instruments of Medicine , Surgery & Emergency department of SMIMER Hospital."

Quotations are to be sent in sealed cover addressed to office of Medical Superintendent, SMIMER Hospital by post/courier so as to reach on Dt. 04 /04 /2025. Quotation received thereafter will not be accepted.

Medical Superintendent
SMIMER Hospital

Price-Bid

Sr. No.	Item Name	Qty.	Department	Unit Rate (Rs.)	Total Amount (Rs.)
1.	<p>Multipara Monitor Model: Comen 80 Sr. No. KE190102019B</p> <p>➤ Compatible reusable Clip Type/ Soft Rubber Adult SpO2 Probe.</p> <ul style="list-style-type: none"> • The spo2 probe must be compitable with the above machines spo2 technology • Cable Lenght: 3mtr. approx • Cable Material : TPU Jacket material & latex free • SpO2 Accuracy: For 70 To 90% saturation ± 3% accuracy. • The bidders are requested to kindly visit the site before bidding to get a complete overview of the unit and its spo2 technology. • Should be ISO/CE/BIS certified Products Only. • Warranty : 1 Year comprehensive 	01	Surgery (2AICU)		
2.	<p>Multipara Monitor Make:Contec Model: CMS 8000 Sr. No. 21030400219</p> <p>➤ Compatible reusable Clip Type/ Soft Rubber Adult SpO2 Probe.</p> <ul style="list-style-type: none"> • The spo2 probe must be compitable to the contec monitors spo2 technology • Cable Lenght: 3mtr. approx • Cable Material : TPU Jacket material & latex free • SpO2 Accuracy: For 99 To 70% saturation ± 2% accuracy. • Should be ISO/CE/BIS certified Products Only. ➤ Warranty : 1 Year comprehensive. 	01	Emergency		
3.	<p>Multipara bedside monitor Make : DRAEGER Model: Vista 120S Sr. no. SISLJ0015,SISLJ0037,SISLJ0046, SISLJ0026, SISLJ0058</p> <ul style="list-style-type: none"> • Replacement of Compatible 5 Lead ECG cable ➤ Compatible reusable 5 ECG limb lead grabber type wires with trunk cable. ➤ Should have grabber type limb wires. ➤ Should have detachable limb wires(preferable) so that each wire can be replaced as and when required. ➤ Leads should be color coded as per IEC ➤ Should be ISO/CE/BIS/FDA certified. One year comprehensive warranty 	05	Medicine (E4AICU-1, E4C Ward-1) Surgery (2BICU/SICU-3)		
4.	<p>Multipara bedside monitor Make : DRAEGER Model: Vista 120S Sr. no.SISLJ0058</p> <p>➤ Compatible reusable Clip Type/ Soft Rubber SpO2 Probe</p> <ul style="list-style-type: none"> • Compatible reusable Spo1 probe must be compitable with above units Spo2 technology • Cable lenth:- 3 meter approx • Cable material:- TPU jacket material & latex free 	01	Surgrey (2BICU/SICU)		

	<ul style="list-style-type: none"> • Spo2 Accuracy:- For 70 to 100% $\pm 2\%$ accuracy • Should be ISO/CE/BIS certified product. • Kindly visit the site for complete overview of the machine/probe and spo2 technology <p>Warranty 1 year Comprehensive Warranty</p>				
5.	<p>Multipara Monitor Make: Comen Model: C- 80 Sr. No.KE190426012B</p> <p>➤ Replacement of compatible Nellcor module and main board</p> <ul style="list-style-type: none"> • Main board PCB : Comen make • The above mentioned both the board must be compatible with the internal PCBs of COMEN C-80 • Should be ISO/CE/BIS certified Products Only. • Warranty : 1 Year comprehensive 	01	Casualty		
6.	<p>Multipara Monitor Make: philips Model:MX-450, Sr.no.DE35195120, DE35195118, DE35195125, DE35195119, DE35195112,DE35195129, DE35195140, DE35195118</p> <p>➤ Replacement of Compatible adult SpO2 Probe.</p> <ul style="list-style-type: none"> • Must be Compatible with reusable Spo2 probe soft rubber/clip type for Philips • Cable Length: 3mtr. approx • Cable Material: TPU jacket material & Latex free. • SpO2 Accuracy: For 70 to 100% $\pm 3\%$ • Should be ISO/CE/BIS Certified Products only. <p>Warranty: One year Comprehensive</p> <p>➤ Replacement of Compatible 5 Lead ECG cable</p> <ul style="list-style-type: none"> • Compatible reusable 5 ECG limb lead grabber type wires with trunk cable. • Should have grabber type limb wires. • Should have detachable limb wires(preferable) so that each wire can be replaced as and when required. • All the lead wires must be encased in a lockable socket(if detachable) • Must be defibrillator shock proof. • Should be ISO/CE/BIS certified. • One year comprehensive warranty 	05	Medicine (MICU 1 & 2)		
		04	Medicine (MICU 1)		

Note: Above mentioned Instrument should be physically checked in, Department Medicine,Surgery,Casualty & Emergency by the Quoting supplier prior to Quoting in SMIMER Hospital.

Medical Superintendent
SMIMER Hospital